

Manufacturer Disclosure Statement for Medical Device Security -- MDS2

Question ID	Question	DUET DRF	See note
DOC-1	Manufacturer Name	CMT (manufactured for Thales)	—
DOC-2	Device Description	DUET DRF	—
DOC-3	Device Model	Software Version 1.20.00	—
DOC-4	Document ID	Chrono # RAD20_1Y087	REV 2
DOC-5	Manufacturer Contact Information	x-rayimaging.support@thalesgroup.com	—
DOC-6	Intended use of device in network-connected environment:	Receive Worklist, Send Images With Patient Data over LAN	In order to connect to Worklist, Printer and PACS
DOC-7	Document Release Date	March 11, 2020	—
DOC-8	Coordinated Vulnerability Disclosure: Does the manufacturer have a vulnerability disclosure program for this device?	Yes	As part of monthly scanning for new vulnerabilities
DOC-9	ISAO: Is the manufacturer part of an Information Sharing and Analysis Organization?	Yes	—
DOC-10	Diagram: Is a network or data flow diagram available that indicates connections to other system components or expected external resources?	Yes	—
DOC-11	SaMD: Is the device Software as a Medical Device (i.e. software-only, no hardware)?	No	—
DOC-11.1	Does the SaMD contain an operating system?	N/A	—
DOC-11.2	Does the SaMD rely on an owner/operator provided operating system?	N/A	—
DOC-11.3	Is the SaMD hosted by the manufacturer?	N/A	—
DOC-11.4	Is the SaMD hosted by the customer?	N/A	—
MANAGEMENT OF PERSONALLY IDENTIFIABLE INFORMATION		Yes, No, N/A, or See Note	Note #
MPII-1	Can this device display, transmit, store, or modify personally identifiable information (e.g. electronic Protected Health Information (ePHI))?	Yes	Full Name, Civil ID, DOB, Gender, BMI, Pregnancy Status, Xray Image__
MPII-2	Does the device maintain personally identifiable information?	Yes	—
MPII-2.1	Does the device maintain personally identifiable information temporarily in volatile memory (i.e., until cleared by power-off or reset)?	Yes	—
MPII-2.2	Does the device store personally identifiable information persistently on internal media?	Yes	—
MPII-2.3	Is personally identifiable information preserved in the device's non-volatile memory until explicitly erased?	Yes	—

MPII-2.4	Does the device store personally identifiable information in a database?	Yes	SQL Localdb storage without external advertising
MPII-2.5	Does the device allow configuration to automatically delete local personally identifiable information after it is stored to a long term solution?	Yes	—
MPII-2.6	Does the device import/export personally identifiable information with other systems (e.g., a wearable monitoring device might export personally identifiable information to a server)?	Yes	DICOM communication
MPII-2.7	Does the device maintain personally identifiable information when powered off, or during power service interruptions?	Yes	—
MPII-2.8	Does the device allow the internal media to be removed by a service technician (e.g., for separate destruction or customer retention)?	Yes	—
MPII-2.9	Does the device allow personally identifiable information records be stored in a separate location from the device's operating system (i.e. secondary internal drive, alternate drive partition, or remote storage location)?	Yes	—
MPII-3	Does the device have mechanisms used for the transmitting, importing/exporting of personally identifiable information?	Yes	—
MPII-3.1	Does the device display personally identifiable information (e.g., video display, etc.)?	Yes	—
MPII-3.2	Does the device generate hardcopy reports or images containing personally identifiable information?	Yes	—
MPII-3.3	Does the device retrieve personally identifiable information from or record personally identifiable information to removable media (e.g., removable-HDD, USB memory, DVD-R/RW, CD-R/RW, tape, CF/SD card, memory stick, etc.)?	Yes	—
MPII-3.4	Does the device transmit/receive or import/export personally identifiable information via dedicated cable connection (e.g., RS-232, RS-423, USB, FireWire, etc.)?	No	—
MPII-3.5	Does the device transmit/receive personally identifiable information via a wired network connection (e.g., RJ45, fiber optic, etc.)?	Yes	—
MPII-3.6	Does the device transmit/receive personally identifiable information via a wireless network connection (e.g., WiFi, Bluetooth, NFC, infrared, cellular, etc.)?	No	—
MPII-3.7	Does the device transmit/receive personally identifiable information over an external network (e.g., Internet)?	No	—
MPII-3.8	Does the device import personally identifiable information via scanning a document?	No	—
MPII-3.9	Does the device transmit/receive personally identifiable information via a proprietary protocol?	No	DICOM, HL7 only
MPII-3.10	Does the device use any other mechanism to transmit, import or export personally identifiable information?	No	—

AUTOMATIC LOGOFF (ALOF)

The device's ability to prevent access and misuse by unauthorized users if device is left idle for a period of time.

ALOF-1	Can the device be configured to force reauthorization of logged-in user(s) after a predetermined length of inactivity (e.g., auto-logoff, session lock, password protected screen saver)?	Yes	—
ALOF-2	Is the length of inactivity time before auto-logoff/screen lock user or administrator configurable?	Yes	0-1000 Minutes(0 infinity),Default 30 minutes if activated

AUDIT CONTROLS (AUDT)

The ability to reliably audit activity on the device.

AUDT-1	Can the medical device create additional audit logs or reports beyond standard operating system logs?	Yes	—
AUDT-1.1	Does the audit log record a USER ID?	Yes	—
AUDT-1.2	Does other personally identifiable information exist in the audit trail?	No	—
AUDT-2	Are events recorded in an audit log? If yes, indicate which of the following events are recorded in the audit log:	Yes	—
AUDT-2.1	Successful login/logout attempts?	Yes	—
AUDT-2.2	Unsuccessful login/logout attempts?	No	—
AUDT-2.3	Modification of user privileges?	No	—
AUDT-2.4	Creation/modification/deletion of users?	No	—
AUDT-2.5	Presentation of clinical or PII data (e.g. display, print)?	Yes	—
AUDT-2.6	Creation/modification/deletion of data?	Yes	There is a log for deleted images ePHI and images, optional to anonymize
AUDT-2.7	Import/export of data from removable media (e.g. USB drive, external hard drive, DVD)?	Yes	—
AUDT-2.8	Receipt/transmission of data or commands over a network or point-to-point connection?	Yes	Internal lan to pacs server dicom ePHI
AUDT-2.8.1	Remote or on-site support?	Yes	Team Viewer available
AUDT-2.8.2	Application Programming Interface (API) and similar activity?	N/A	—
AUDT-2.9	Emergency access?	Yes	Emergency log,by analyzing time on logtech can determine what action done
AUDT-2.10	Other events (e.g., software updates)?	Yes	Delete images
AUDT-2.11	Is the audit capability documented in more detail?	No	—
AUDT-3	Can the owner/operator define or select which events are recorded in the audit log?	No	—
AUDT-4	Is a list of data attributes that are captured in the audit log for an event available?	Yes	Date,Module,Line,Activity
AUDT-4.1	Does the audit log record date/time?	Yes	—
AUDT-4.1.1	Can date and time be synchronized by Network Time Protocol (NTP) or equivalent time source?	Yes	—
AUDT-5	Can audit log content be exported?	Yes	—
AUDT-5.1	Via physical media?	Yes	Manually with technician access
AUDT-5.2	Via IHE Audit Trail and Node Authentication (ATNA) profile to SIEM?	No	—
AUDT-5.3	Via Other communications (e.g., external service device, mobile applications)?	No	—

AUDT-5.4	Are audit logs encrypted in transit or on storage media?	Yes	encrypt on transit
AUDT-6	Can audit logs be monitored/reviewed by owner/operator?	No	On rest as part of bitlocker
AUDT-7	Are audit logs protected from modification?	Yes	Can be reviewed by Field Engineer
AUDT-7.1	Are audit logs protected from access?	Yes	no access to log using non-admin
AUDT-8	Can audit logs be analyzed by the device?	No	personal
			no access to log using non-admin
			personal
			—

AUTHORIZATION (AUTH)

The ability of the device to determine the authorization of users.

AUTH-1	Does the device prevent access to unauthorized users through user login requirements or other mechanism?	Yes	passwords
AUTH-1.1	Can the device be configured to use federated credentials management of users for authorization (e.g., LDAP, OAuth)?	Yes	—
AUTH-1.2	Can the customer push group policies to the device (e.g., Active Directory)?	No	—
AUTH-1.3	Are any special groups, organizational units, or group policies required?	No	—
AUTH-2	Can users be assigned different privilege levels based on 'role' (e.g., user, administrator, and/or service, etc.)?	Yes	—
AUTH-3	Can the device owner/operator grant themselves unrestricted administrative privileges (e.g., access operating system or application via local root or administrator account)?	Yes	OEM can cancel this option on site
AUTH-4	Does the device authorize or control all API access requests?	N/A	
AUTH-5	Does the device run in a restricted access mode, or 'kiosk mode', by default?	Yes	Kiosk mode load directly software as shell

CYBER SECURITY PRODUCT UPGRADES (CSUP)

The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade device's security patches.

CSUP-1	Does the device contain any software or firmware which may require security updates during its operational life, either from the device manufacturer or from a third-party manufacturer of the software/firmware? If no, answer "N/A" to questions in this section.	Yes	—
CSUP-2	Does the device contain an Operating System? If yes, complete 2.1-2.4.	Yes	—
CSUP-2.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	Yes	CMT issues FTU for each software update
CSUP-2.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	No	Any patch or SW installation must be backed up by a written authorization from the vendor

CSUP-2.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-2.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	No	—
CSUP-3	Does the device contain Drivers and Firmware? If yes, complete 3.1-3.4.	Yes	—
CSUP-3.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	No	—
CSUP-3.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	—
CSUP-3.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-3.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	No	—
CSUP-4	Does the device contain Anti-Malware Software? If yes, complete 4.1-4.4.	Yes	Mcafee application control
CSUP-4.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	No	—
CSUP-4.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	—
CSUP-4.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-4.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	No	—
CSUP-5	Does the device contain Non-Operating System commercial off-the-shelf components? If yes, complete 5.1-5.4.	Yes	—
CSUP-5.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	No	—
CSUP-5.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	—
CSUP-5.3	Does the device have the capability to receive remote installation of patches or software updates?	No	—
CSUP-5.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	No	—
CSUP-6	Does the device contain other software components (e.g., asset management software, license management)? If yes, please provide details or reference in notes and complete 6.1-6.4.	No	—
CSUP-6.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	N/A	—

CSUP-6.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A	—
CSUP-6.3	Does the device have the capability to receive remote installation of patches or software updates?	N/A	—
CSUP-6.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	N/A	—
CSUP-7	Does the manufacturer notify the customer when updates are approved for installation?	Yes	Manufacture to OEM notification as part of new version
CSUP-8	Does the device perform automatic installation of software updates?	No	—
CSUP-9	Does the manufacturer have an approved list of third-party software that can be installed on the device?	N/A	—
CSUP-10	Can the owner/operator install manufacturer-approved third-party software on the device themselves?	No	—
CSUP-10.1	Does the system have mechanism in place to prevent installation of unapproved software?	Yes	Mcafee application control
CSUP-11	Does the manufacturer have a process in place to assess device vulnerabilities and updates?	Yes	Monthly scan for vulnerabilities and release mitigation if required in new versions
CSUP-11.1	Does the manufacturer provide customers with review and approval status of updates?	Yes	Per request
CSUP-11.2	Is there an update review cycle for the device?	Yes	Every SW version release

HEALTH DATA DE-IDENTIFICATION (DIDT)

The ability of the device to directly remove information that allows identification of a person.

DIDT-1	Does the device provide an integral capability to de-identify personally identifiable information?	Yes	
DIDT-1.1	Does the device support de-identification profiles that comply with the DICOM standard for de-identification?	No	—

DATA BACKUP AND DISASTER RECOVERY (DTBK)

The ability to recover after damage or destruction of device data, hardware, software, or site configuration information.

DTBK-1	Does the device maintain long term primary storage of personally identifiable information / patient information (e.g. PACS)?	No	—
DTBK-2	Does the device have a "factory reset" function to restore the original device settings as provided by the manufacturer?	Yes	—

DTBK-3	Does the device have an integral data backup capability to removable media?	Yes	Backup of database and logs to analyze issues,can be anonymized
DTBK-4	Does the device have an integral data backup capability to remote storage?	No	
DTBK-5	Does the device have a backup capability for system configuration information, patch restoration, and software restoration?	Yes	
DTBK-6	Does the device provide the capability to check the integrity and authenticity of a backup?	Yes	Checksum exist

EMERGENCY ACCESS (EMRG)

The ability of the device user to access personally identifiable information in case of a medical emergency situation that requires immediate access to stored personally identifiable information.

EMRG-1	Does the device incorporate an emergency access (i.e. “break-glass”) feature?	Yes	Restricted Access to only emergency patient.
--------	---	-----	--

HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU)

How the device ensures that the stored data on the device has not been altered or destroyed in a non-authorized manner and is from the originator.

IGAU-1	Does the device provide data integrity checking mechanisms of stored health data (e.g., hash or digital signature)?	No	—
IGAU-2	Does the device provide error/failure protection and recovery mechanisms for stored health data (e.g., RAID-5)?	Yes	RAID 1 Mirror

MALWARE DETECTION/PROTECTION (MLDP)

The ability of the device to effectively prevent, detect and remove malicious software (malware).

MLDP-1	Is the device capable of hosting executable software?	Yes	only provided by manufacturer McAfee application control, Windows Defender McAfee application control, Windows Defender The system includes a default anti-malware so this is N/A
MLDP-2	Does the device support the use of anti-malware software (or other anti-malware mechanism)? Provide details or reference in notes.	Yes	
MLDP-2.1	Does the device include anti-malware software by default?	Yes	
MLDP-2.2	Does the device have anti-malware software available as an option?	N/A	— Changes could be consodered per request
MLDP-2.3	Does the device documentation allow the owner/operator to install or update anti-malware software?	No	
MLDP-2.4	Can the device owner/operator independently (re-)configure anti-malware settings?	No	

MLDP-2.5	Does notification of malware detection occur in the device user interface?	No	
MLDP-2.6	Can only manufacturer-authorized persons repair systems when malware has been detected?	No	
MLDP-2.7	Are malware notifications written to a log?	Yes	Windows Log
MLDP-2.8	Are there any restrictions on anti-malware (e.g., purchase, installation, configuration, scheduling)?	Yes	
MLDP-3	If the answer to MLDP-2 is NO, and anti-malware cannot be installed on the device, are other compensating controls in place or available?	N/A	—
MLDP-4	Does the device employ application whitelisting that restricts the software and services that are permitted to be run on the device?	Yes	Could be considered per request
MLDP-5	Does the device employ a host-based intrusion detection/prevention system?	No	—
MLDP-5.1	Can the host-based intrusion detection/prevention system be configured by the customer?	No	—
MLDP-5.2	Can a host-based intrusion detection/prevention system be installed by the customer?	No	—

NODE AUTHENTICATION (NAUT)

The ability of the device to authenticate communication partners/nodes.

NAUT-1	Does the device provide/support any means of node authentication that assures both the sender and the recipient of data are known to each other and are authorized to receive transferred information (e.g. Web APIs, SMTP, SNMP)?	Yes	DICOM AE
NAUT-2	Are network access control mechanisms supported (E.g., does the device have an internal firewall, or use a network connection white list)?	Yes	Windows Firewall Technician can change it according users requests.
NAUT-2.1	Is the firewall ruleset documented and available for review?	Yes	
NAUT-3	Does the device use certificate-based network connection authentication?	No	—

CONNECTIVITY CAPABILITIES (CONN)

All network and removable media connections must be considered in determining appropriate security controls. This section lists connectivity capabilities that may be present on the device.

CONN-1	Does the device have hardware connectivity capabilities?	Yes	Network,USB Mass storage
CONN-1.1	Does the device support wireless connections?	Yes	For Wireless detector Optional for wireless detector WPA2-PSK(by HARDWARE AP)
CONN-1.1.1	Does the device support Wi-Fi?	Yes	
CONN-1.1.2	Does the device support Bluetooth?	No	—
CONN-1.1.3	Does the device support other wireless network connectivity (e.g. LTE, Zigbee, proprietary)?	No	—
CONN-1.1.4	Does the device support other wireless connections (e.g., custom RF controls, wireless detectors)?	Yes	Optional Infrared to configure wireless detector

CONN-1.2	Does the device support physical connections?	Yes	—
CONN-1.2.1	Does the device have available RJ45 Ethernet ports?	Yes	—
CONN-1.2.2	Does the device have available USB ports?	Yes	—
CONN-1.2.3	Does the device require, use, or support removable memory devices?	Yes	For Capture service data for analyze or backup/restore configuraion
CONN-1.2.4	Does the device support other physical connectivity?	Yes	Serial ports to generator, keypad
CONN-2	Does the manufacturer provide a list of network ports and protocols that are used or may be used on the device?	Yes	—
CONN-3	Can the device communicate with other systems within the customer environment?	Yes	—
CONN-4	Can the device communicate with other systems external to the customer environment (e.g., a service host)?	No	—
CONN-5	Does the device make or receive API calls?	No	—
CONN-6	Does the device require an internet connection for its intended use?	No	—
CONN-7	Does the device support Transport Layer Security (TLS)?	No	—
CONN-7.1	Is TLS configurable?	N/A	—
CONN-8	Does the device provide operator control functionality from a separate device (e.g., telemedicine)?	No	—

PERSON AUTHENTICATION (PAUT)

The ability to configure the device to authenticate users.

PAUT-1	Does the device support and enforce unique IDs and passwords for all users and roles (including service accounts)?	Yes	—
PAUT-1.1	Does the device enforce authentication of unique IDs and passwords for all users and roles (including service accounts)?	Yes	—
PAUT-2	Is the device configurable to authenticate users through an external authentication service (e.g., MS Active Directory, NDS, LDAP, OAuth, etc.)?	Yes	Optional Software only authentication or Active Directory
PAUT-3	Is the device configurable to lock out a user after a certain number of unsuccessful logon attempts?	See Notes	Only when using Windows Authentication
PAUT-4	Are all default accounts (e.g., technician service accounts, administrator accounts) listed in the documentation?	Yes	—
PAUT-5	Can all passwords be changed?	Yes	—
PAUT-6	Is the device configurable to enforce creation of user account passwords that meet established (organization specific) complexity rules?	Yes	—
PAUT-7	Does the device support account passwords that expire periodically?	See Notes	Can be configured via Windows settings
PAUT-8	Does the device support multi-factor authentication?	No	—
PAUT-9	Does the device support single sign-on (SSO)?	No	—
PAUT-10	Can user accounts be disabled/locked on the device?	Yes	Windows account can be locked
PAUT-11	Does the device support biometric controls?	No	—

PAUT-12	Does the device support physical tokens (e.g. badge access)?	No	—
PAUT-13	Does the device support group authentication (e.g. hospital teams)?	No	—
PAUT-14	Does the application or device store or manage authentication credentials?	Yes	On database
PAUT-14.1	Are credentials stored using a secure method?	Yes	—

PHYSICAL LOCKS (PLOK)

Physical locks can prevent unauthorized users with physical access to the device from compromising the integrity and confidentiality of personally identifiable information stored on the device or on removable media

PLOK-1	Is the device software only? If yes, answer “N/A” to remaining questions in this section.	No	—
PLOK-2	Are all device components maintaining personally identifiable information (other than removable media) physically secure (i.e., cannot remove without tools)?	Yes	—
PLOK-3	Are all device components maintaining personally identifiable information (other than removable media) physically secured behind an individually keyed locking device?	Yes	—
PLOK-4	Does the device have an option for the customer to attach a physical lock to restrict access to removable media?	Yes	—

ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE CYCLE (RDMP)

Manufacturer’s plans for security support of third-party components within the device’s life cycle.

RDMP-1	Was a secure software development process, such as ISO/IEC 27034 or IEC 62304, followed during product development?	Yes	IEC 62304, IEC 27799 followed
RDMP-2	Does the manufacturer evaluate third-party applications and software components included in the device for secure development practices?	Yes	—
RDMP-3	Does the manufacturer maintain a web page or other source of information on software support dates and updates?	No	—
RDMP-4	Does the manufacturer have a plan for managing third-party component end-of-life?	No	—

SOFTWARE BILL OF MATERIALS (SBoM)

A Software Bill of Material (SBoM) lists all the software components that are incorporated into the device being described for the purpose of operational security planning by the healthcare delivery organization. This section supports controls in the RDMP section.

SBOM-1	Is the SBoM for this product available?	Yes	
SBOM-2	Does the SBoM follow a standard or common method in describing software components?	Yes	IEC 62304 compliant
SBOM-2.1	Are the software components identified?	Yes	—
SBOM-2.2	Are the developers/manufacturers of the software components identified?	Yes	—

SBOM-2.3	Are the major version numbers of the software components identified?	Yes	—
SBOM-2.4	Are any additional descriptive elements identified?	Yes	—
SBOM-3	Does the device include a command or process method available to generate a list of software components installed on the device?	No	—
SBOM-4	Is there an update process for the SBOM?	Yes	IEC 62304 checklist

SYSTEM AND APPLICATION HARDENING (SAHD)

The device's inherent resistance to cyber attacks and malware.

SAHD-1	Is the device hardened in accordance with any industry standards?	Yes	DOD STIG Levels
SAHD-2	Has the device received any cybersecurity certifications?	See Notes	The manufacture Facility receive
SAHD-3	Does the device employ any mechanisms for software integrity checking	Yes	ISO27001 Certification
SAHD-3.1	Does the device employ any mechanism (e.g., release-specific hash key, checksums, digital signature, etc.) to ensure the installed software is manufacturer-authorized?	Yes	—
SAHD-3.2	Does the device employ any mechanism (e.g., release-specific hash key, checksums, digital signature, etc.) to ensure the software updates are the manufacturer-authorized updates?	Yes	Hashing on software version
SAHD-4	Can the owner/operator perform software integrity checks (i.e., verify that the system has not been modified or tampered with)?	Yes	Hashing on softwareupdate version
SAHD-5	Is the system configurable to allow the implementation of file-level, patient level, or other types of access controls?	No	—
SAHD-5.1	Does the device provide role-based access controls?	No	—
SAHD-6	Are any system or user accounts restricted or disabled by the manufacturer at system delivery?	Yes	Administrator user is disabled,
SAHD-6.1	Are any system or user accounts configurable by the end user after initial configuration?	Yes	Guest user is renamed and disabled
SAHD-6.2	Does this include restricting certain system or user accounts, such as service technicians, to least privileged access?	Yes	—
SAHD-7	Are all shared resources (e.g., file shares) which are not required for the intended use of the device disabled?	Yes	By LOR software settings
SAHD-8	Are all communication ports and protocols that are not required for the intended use of the device disabled?	Yes	—
SAHD-9	Are all services (e.g., telnet, file transfer protocol [FTP], internet information server [IIS], etc.), which are not required for the intended use of the device deleted/disabled?	Yes	—
SAHD-10	Are all applications (COTS applications as well as OS-included applications, e.g., MS Internet Explorer, etc.) which are not required for the intended use of the device deleted/disabled?	Yes	—
SAHD-11	Can the device prohibit boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)?	Yes	—

SAHD-12	Can unauthorized software or hardware be installed on the device without the use of physical tools?	No	Hardware by screwdriver, software cannot be installed without field engineer
SAHD-13	Does the product documentation include information on operational network security scanning by users?	No	—
SAHD-14	Can the device be hardened beyond the default provided state?	Yes	We can consider hardening requests, such as registry changes, block services, etc.
SAHD-14.1	Are instructions available from vendor for increased hardening?	No	
SHAD-15	Can the system prevent access to BIOS or other bootloaders during boot?	Yes	Password Access
SAHD-16	Have additional hardening methods not included in 2.3.19 been used to harden the device?	No	—

SECURITY GUIDANCE (SGUD)

Availability of security guidance for operator and administrator of the device and manufacturer sales and service.

SGUD-1	Does the device include security documentation for the owner/operator?	Yes	—
SGUD-2	Does the device have the capability, and provide instructions, for the permanent deletion of data from the device or media?	Yes	—
SGUD-3	Are all access accounts documented?	Yes	—
SGUD-3.1	Can the owner/operator manage password control for all accounts?	Yes	—
SGUD-4	Does the product include documentation on recommended compensating controls for the device?	No	—

HEALTH DATA STORAGE CONFIDENTIALITY (STCF)

The ability of the device to ensure unauthorized access does not compromise the integrity and confidentiality of personally identifiable information stored on the device or removable media.

STCF-1	Can the device encrypt data at rest?	Yes	—
STCF-1.1	Is all data encrypted or otherwise protected?	Yes	
STCF-1.2	Is the data encryption capability configured by default?	Yes	
STCF-1.3	Are instructions available to the customer to configure encryption?	Yes	For Technician User Only
STCF-2	Can the encryption keys be changed or configured?	Yes	—
STCF-3	Is the data stored in a database located on the device?	Yes	SQL local database internally
STCF-4	Is the data stored in a database external to the device?	No	—

TRANSMISSION CONFIDENTIALITY (TXCF)

The ability of the device to ensure the confidentiality of transmitted personally identifiable information.

TXCF-1	Can personally identifiable information be transmitted only via a point-to-point dedicated cable?	No	—
TXCF-2	Is personally identifiable information encrypted prior to transmission via a network or removable media?	Yes	captured data is always encrypted
TXCF-2.1	If data is not encrypted by default, can the customer configure encryption options?	No	—
TXCF-3	Is personally identifiable information transmission restricted to a fixed list of network destinations?	Yes	—
TXCF-4	Are connections limited to authenticated systems?	Yes	DICOM AE
TXCF-5	Are secure transmission methods supported/implemented (DICOM, HL7, IEEE 11073)?	No	—

TRANSMISSION INTEGRITY (TXIG)

The ability of the device to ensure the integrity of transmitted data.

TXIG-1	Does the device support any mechanism (e.g., digital signatures) intended to ensure data is not modified during transmission?	No	—
TXIG-2	Does the device include multiple sub-components connected by external cables?	Yes	Serial connection to generator, CAN bus connection to Remote table

REMOTE SERVICE (RMOT)			
<i>Remote service refers to all kinds of device maintenance activities performed by a service person via network or other remote connection.</i>			
RMOT-1	Does the device permit remote service connections for device analysis or repair?	Yes	—
RMOT-1.1	Does the device allow the owner/operator to initiative remote service sessions for device analysis or repair?	Yes	—
RMOT-1.2	Is there an indicator for an enabled and active remote session?	See Notes	Indication appears when using high privilege sessions
RMOT-1.3	Can patient data be accessed or viewed from the device during the remote session?	Yes	—
RMOT-2	Does the device permit or use remote service connections for predictive maintenance data?	No	—
RMOT-3	Does the device have any other remotely accessible functionality (e.g. software updates, remote training)?	No	—

OTHER SECURITY CONSIDERATIONS (OTHR)

NONE